



TO: USERS OF POWERED AIR PURIFYING RESPIRATORS USING IMMEDIATE RESPONSE TECHNOLOGIES MANUFACTURED C420 BLOWER PART NUMBERS M-60031-001 and 4056916

AirBoss Defense has identified a potential issue with the Immediate Response Technologies (IRT) manufactured C420 Powered Air Purifying (PAPR) blower, part number M-60031-001 and 4056916 manufactured from 2009 to 2011. These blowers are beyond our standard warranty period and the purpose of this correspondence is to give users notice of how they can identify if they have an affected blower.

Blowers manufactured from 2009 to 2011 may experience low or unassisted airflow; this condition may be accompanied by an unusual noise from the blower. This condition is identifiable during normal pre-checks as described in the User Instructions before use and during use of the C420 PAPR audibly via unusual noise or via a noticeable decrease in airflow to the face piece.

Blowers manufactured outside of the 2009 to 2011 date range are not affected by this notice.


IDENTIFICATION OF AFFECTED C420 BLOWERS

To determine if your blower is affected by this user notice please review the list below for applicable serial numbers and compare to the serial number located on the back face of the C420 blower in the center. This notice only affects part numbers M-60031-001 and 4056916 with Serial Numbers: 13049 to 20712.

If you possess a blower exhibiting these conditions AirBoss Defense recommends you discontinue use.



USER NOTICE



**Part Number
and Serial Number**

NOTIFICATION:

Please contact AirBoss Defense Sales department at **1-855-453-7137**, or via email at **Quality@airbossdefense.com** for further information regarding this User Notification.